

- Measure (3 ranges) up to 40V, 400mA, 40kΩ
- Output voltage (3 ranges) to 10V, 1µV resolution
- Output current (3 ranges) to 20mA, 1µA resolution
- Output resistance: 0 to 400Ω , 0.01Ω resolution
- Thermocouple measure/simulate: J, K, T, E, R, S, B, N
- RTD measure and simulate: PT100
- Ramp function
- Battery or mains operation
- · Housed in robust field case

DESCRIPTION

The 7050 Process Calibrator combines the essential functions of measurement and simulation of volts, millivolts, milliamps, and ohms with the direct readout simulation of thermocouples and RTD's in °C or °F units.

The instrument is microprocessor based and housed in a rugged case designed for use in the field. Input and Output functions are displayed simultaneously and the switch controls allow for easy and precise operation.

Five calibration points are instantaneously accessible through a preset percentage selector and all functions may be generated in a fully programmable ramp format.

In addition to the standard calibration facilities the instrument may be used as a signal converter. Any electronic input signal can be converted and output as a proportional signal of 4-20mA, 0-5V or sq. root 4-20mA.

The 7050 is highly adaptable, IPT 68 and ITS 90 curves are accommodated and non-standard thermocouples and RTD's may be programmed on request.

Specifations – Voltage / Current / Resistance Generation and Measure

INPUT

Voltage

Range Resolution		Accuracy	
40V	1mV	0.03%	
4V	100μV	0.03%	
400mV	10μV	0.03%	
40mV	1 <i>μ</i> V	0.03%	

Current

Range	Resolution	Accuracy
400mA	10μA	0.03%
40mA	1μΑ	0.03%
4-20mA/0-100%	0.01%	0.06%

Resistance

Range	Resolution	Accuracy
40000Ω	1Ω	0.03%
4000Ω	0.1Ω	0.03%
400Ω	0.01Ω	0.06%

OUTPUT

Voltage

Range	Resolution	Accuracy	
-2 to +10V	1mV	0.06%	
-100mV to +400mV	10μV	0.03%	
-10mV to +40mV 1μV 0.03%			
Output Impedance: < 10Ω			

Current

Range	Resolution	Accuracy		
0 to 20mA	10μA	0.03%		
4 to 20mA	1μA	0.03%		
TX Sim (4 to 20mA) 0.01% 0.03%				
Max Load Resistance: 900Ω at 20mA, Open circuit voltage 18- 27V				

Resistance

Range	Resolution	Accuracy
0 to 400Ω	0.01Ω	± 0.03%
Excitation Current 0.2 to 2mA		

Other

Loop drive supply: Nominal 24V, current limit set to 30mA

7050 Specifications

TECHNICAL SPECIFICATIONS - PROCESS SIGNAL & TEMPERATURE MEASUREMENT & SIMULATION

TEMPERATURE

	°C			°F	
Thermocouple Type	Temp Range °C	Accuracy °C	Thermocouple Type	Temp Range °F	Accuracy °F
J	-210 to 0 0 to 710 710 to 1200	± 0.4 ± 0.2 ± 0.9	J	-410 to 32 32 to 1310 1310 to 2192	± 0.7 ± 0.4 ± 1.0
К	-270 to -240 -240 to -180 -180 to -50 -50 to 960 960 to 1370	± 3.0 ± 1.2 ± 0.5 ± 0.3 ± 1.7	К	-454 to -430 -430 to -292 -292 to -58 -58 to 1760 1760 to 2498	± 5.4 ± 2.0 ± 0.9 ± 0.6 ± 1.5
Т	-270 to -250 -250 to -180 -180 to -100 -100 to 400	± 3.0 ± 1.0 ± 0.5 ± 0.2	т	-454 to -418 -418 to -292 -292 to -148 -148 to 752	± 5.4 ± 1.8 ± 0.9 ± 0.4
R	-50 to 0 0 to 70 70 to 400 400 to 1000 1000 to 1760	± 3.0 ± 2.0 ± 1.5 ± 1.0 ± 0.8	R	-58 to 32 32 to 158 158 to 752 752 to 1832 1832 to 3200	± 5.4 ± 3.6 ± 2.7 ± 1.8 ± 1.4
S	-50 to 0 0 to 100 100 to 450 450 to 1760	± 2.5 ± 1.8 ± 1.3 ± 0.9	S	-58 to 32 32 to 212 212 to 842 842 to 3200	± 4.5 ± 3.2 ± 2.3 ± 1.6
В	200 to 400 400 to 1000 1000 to 1820	± 5.0 ± 2.0 ± 1.0	В	392 to 752 752 to 1832 1832 to 3308	± 9.0 ± 3.6 ± 1.8
N	0 to 330 330 to 1100 1100 to 1300	± 0.4 ± 0.3 ± 1.5	N	32 to 626 626 to 2012 2012 to 2372	± 0.7 ± 0.5 ± 2.7
E	-230 to 0 0 to 540 540 to 1000	± 0.5 ± 0.2 ± 1.5	E	-382 to 32 32 to 1004 1004 to 1832	± 0.9 ± 0.4 ± 2.7

RTD

PT100 (°C)		P	T100 (°F)
Range	Accuracy	Range	Accuracy
-200 to 850	± 0.3	-392 to1562	± 1.0

Pt100 Linearisation to BS1904 (1984).DIN43760 (1980). Linearisation Acc: 0.1% of reading for Class A, 38.5Ω FI

Cold Junction Compensation...... External (Pt100), Internal, Ice point. Accuracy +/-0.2°C at 23°C

Ramp Fully programmable continuous and stepped

Transmitter Tx function 4-20mA

Signal Converter Any input to any output

GENERAL SPECIFICATION

Battery Rechargeable NiMHd - Internal Charger Displays LCD Dot Matrix, 1 line by 20 charx12mm, alphanumeric.

ORDERING INFORMATION

7050	Process and Thermocouple Calibrator
C180	Factory (NPL Traceable) Calibration Certificate
C192	UKAS Calibration Certificate (ISO 17025)